

شركة مصافي الجنوب/ شركه عامة

WWW.SRC.GOV.IQ

SOUTH REFINERIES COMPANY IS ANNOUNCING THE ISSUANCE OF HEREUNDER REQUISITION . ACCORDINGLY WHOSOEVER WOULD LIKE TO PARTICIPATE AND OFFER SHOULD ATTEND SRC / HEADQUARTER LOCATED IN BASRAH/ SHUAIBA.

GAINING TENDER SPECIFICATIONS FOR AN AMOUNT MENTIONED AGAINST EACH TENDER CASH IN ADVANCE IT IS NON REFUNDABLE WHATEVER REASONS AND WHO WINS THE TENDER SHALL PAY THE COSTS OF LAST ANNOUNCEMENT.THE WELL KNOWN INTERNATIONAL COMPANIES WHICH HAVE NO BRANCH OR REPRESENTATIVE IN IRAQ CAN PARTICIPATE AND SEND THEIR OFFERS THROUGH DHL OR FEDEX OR ANY OTHER WAY WITHIN TIME DECIDED WITHOUT ATTACHING THE BILL OF PURCHASING BIDDING DOCUMENTS.

CLOSING DATE WILL BE UNTIL THE END OF OFFICIAL WORKING HOURES OF SUNDAY 28 / APRIL /2013.

ITEM	REQ NO	DESCRIPTION	PRICE
1-	58 /2013/ E	Update apparatus for Determination of Aromatic Hydrocarbon Types in Aviation Fuel and petroleum Distillates by High performance Liquid Chromatography (HPLC) according to ASTM D6379	\$100

Please, see attached conditions

FOR/ DIRECTOR GENERAL
Akeel SH. Ahmed

24 3 / 2013

Item	Description	Qty
1	Update apparatus for Determination of Aromatic Hydrocarbon Types in Aviation Fuel and petroleum Distillates by High performance Liquid Chromatography (HPLC) according to ASTM D6379 . <u>Specification:</u> 1- Update apparatus for determination of total Aromatics in Jet A-1 according to ASTM D 6379. 2- Electricity: 220 V/ 50-60 Hz. <u>General requirements:</u> 1- All materials, chemicals and accessories needed for operation of apparatus for period (2) Years .Number of yearly tests 200. 2-A calibration materials conformity with ASTM D 6379 for period (2) years. a)System Resolution Standards *Cat No. D-6379-SRS-PAK *Cat No. D-6379-SRS-R1-PAK b) Calibration Curves: *Cat No. D-6379-SET-PAK / 5 x (4 x 1 mL) *Cat No. D-6379-10X-SET-PAK 5 x (4 x 1 mL) 3- A list of spare parts , glassware and electronics needed with its part number for operation and maintenance of apparatus for a period of (3) years, 4- A copy of maintenance and service manual book in English Including electronic maintenance sheets and electronic calibration, adjustment methods . 5- A suitable connection with PC, printer and network. 6- The latest issue of test method ASTM D6379. 8- An apparatus calibration certificate. 9- The guarantee of the apparatus for one year operation. ∑ Training for two person in country of Origin of the apparatus least 2 weeks without travel days. ∑ Submit all tools needed for maintenance if present . ∑ Mention the country of origin of the apparatus. Not : installation and operation must be in our lap rotary. ∑ Contain the origin of the apparatus that must be USA, west Europe or Japan.	1 5 x 1 mL 5 x 1 mL 5 x (4x1mL) 5 x (4x1mL)